

Instrument	TOC/DU	Time interval/UTC	Difference to BTS/%	SD of values used/DU
BTS	267.5	09:00–11:00	–	1.0
QASUME	271.7	09:00–11:00	+1.6	0.5
IZO Brewer	269.7	09:00–11:00	+0.8	0.4
NOAA Dobson	265.5	09:00–11:00	–0.8	0.6
BTS	270	12:30–13:00	–	0.7
Aura OMI	275	12:30–13:00	+1.8	–