

Sensor	Measurement %	Sensor on and not logging %	Completely missing %	Comments
Aeroqual	82 %	0 %	18 %	45 % of logged values were 0
	73 %	0 %	27 %	42 % of logged values were 0
	81 %	5 %	13 %	32 % of logged values were 0
	87 %	0 %	13 %	
AirAssure	87 %	0 %	13 %	
	87 %	0 %	13 %	
	74 %	0 %	25 %	
AirBeam	62 %	6 %	32 %	
	62 %	6 %	32 %	
	29 %	53 %	18 %	56 % of logged values were 255*
CairClip	63 %	13 %	24 %	No data before 10 Aug 2015
	63 %	23 %	13 %	
	82 %	0 %	18 %	
Dylos	82 %	0 %	18 %	
	72 %	1 %	27 %	
	77 %	0 %	23 %	
OPC-N2	76 %	0 %	24 %	59 % of logged values were 0
	71 %	0 %	29 %	
	82 %	0 %	18 %	
Shinyei	73 %	0 %	27 %	
	87 %	0 %	13 %	
	92 %	0 %	8 %	
Speck	93 %	0 %	7 %	
	96 %	0 %	4 %	
	61 %	0 %	39 %	
TZO A	47 %	0 %	53 %	
	47 %	0 %	53 %	