

Performance characteristics		1 h	6 h	12 h	24 h
Adjustment	E-BAM	E-BAM, T	E-BAM	E-BAM	E-BAM
β_0	3.7 (3.2–4.1)	4.5 (4.1–5.1)	–1.9 (–2.3 to –1.4)	–2.4 (–2.8 to –1.8)	–4.2 (–4.6 to –3.6)
β_1	0.7 (0.6–0.8)	0.7 (0.7–0.8)	1.4 (1.2–1.5)	1.4 (1.3–1.5)	1.6 (1.5–1.8)
β_2	–	–0.06 (–0.07 to –0.05)	–	–	–
R^2	0.40 (0.36–0.41)	0.41 (0.36–0.42)	0.80 (0.77–0.82)	0.84 (0.81–0.86)	0.93 (0.90–0.94)
Mean of ratios ¹	0.97 (0.96–0.97)	0.90 (0.90–0.91)	1.05 (1.04–1.06)	1.01 (1.01–1.02)	1 (1–1.01)
Error ²	201 % (195 %–223 %)	207 % (201 %–229 %)	53 % (50 %–55 %)	35 % (33 %–39 %)	15 % (13 %–18 %)